



دائرة اللوازم والمشتريات

عطاء رقم (T35 /2018.19)

توفير اجهزة مخبرية

2018, 2019



عطاء توفير اجهزة مخبرية

وثائق العطاء:

أ- الجزء الأول:

(1) دعوة العطاء

(2) الشروط والتعليمات التنظيمية للعطاء

(3) طريقة الدفع

ب- الجزء الثاني:

(1) جدول الكميات ومواصفات الفنية



الجزء الأول (1)

إعلان طرح عطاء رقم (T35/2018.19)

توفير أجهزة مخبرية

تدعو الجامعة العربية الأمريكية الشركات المختصة الى المشاركة في العطاء المذكور أعلاه. يمكن الاستفسار أو الحصول على وثائق العطاء من دائرة اللوازم والمشتريات في الجامعة/ مبنى الدوائر الإدارية الطابق الثاني، هاتف- 04 2418888- تحويلة 1488 فاكس 04 2510972 بريد الكتروني pnp@aaup.edu مقابل مبلغ غير مسترد مقداره (50) دولار تدفع في إحدى البنوك المعتمدة وذلك اعتباراً من يوم الاثنين الموافق 2019/6/10.

ملاحظات :

1. تقديم عرضين: فني ومالي، وسيتم دراسة العروض فنياً ومالياً لاختيار العرض المناسب.
2. آخر موعد لتسليم العطاءات هو في تمام الساعة الثانية من يوم الاثنين 2019/6/24 ولنفس المكان.
3. يجب تقديم كفالة دخول عطاء 5% من قيمة العطاء على شكل كفالة بنكية أو شيك بنكي مصدق لصالح الجامعة العربية الأمريكية.
4. الأسعار بالدولار وتشمل جميع الضرائب بما فيها ضريبة القيمة المضافة وعلى المورد تقديم الفواتير الضريبية وشهادة خصم المصدر.
5. الجامعة غير ملزمة بأقل الأسعار وبدون إبداء الأسباب.
6. رسوم الاعلان على من يرسو عليه العطاء.



الشروط والتعليمات التنظيمية للعطاء

(2)

1. على جميع المشاركين في العطاء الالتزام التام بهذه الشروط والتعليمات، وهي تعتبر جزءاً لا يتجزأ من أي أمر شراء أو عقد يبرم مع المشارك الفائز ما لم ينص صراحة على خلاف ذلك في أمر الشراء أو العقد.
2. في هذه الشروط والتعليمات يرمز إلى "الجامعة العربية الامريكية بالاختصار (AAUP).
3. يجب أن تكون الشركة المتقدمة للعطاء مسجلة رسمياً ومشتغلاً مرخصاً.
4. تقدم الأسعار بالدولار الامريكي شاملاً لجميع الضرائب بما في ذلك ضريبة القيمة المضافة (VAT).
5. يلتزم المشارك الفائز بتقديم شهادات خصم المصدر والفواتير الضريبية اللازمة وأية مستندات قانونية أخرى تغطي عملية الشراء.
6. يجب أن تشمل الأسعار على جميع المصاريف المطلوبة من النقل والتركيب والتشغيل والفحص والصيانة والتدريب في المواقع المحددة في جدول المواصفات والكميات المرفق.
7. يجب أن تكون الأسعار المقدمة سارية المفعول لمدة لا تقل عن (90) يوماً من تاريخ تقديم العرض.
8. على المشارك الفائز تقديم كفالة حسن تنفيذ خلال أسبوع من تاريخ الاتفاقية بحيث تعادل (10%) من قيمة الاتفاقية على شكل كفالة بنكية صادرة عن إحدى البنوك العاملة في فلسطين أو شيك مصدق صادر لصالح "الجامعة العربية الامريكية".
9. إذا تخلف المناقص الفائز عن تقديم كفالة حسن التنفيذ عن الموعد المحدد في البند السابق فإنه يحق لـ (AAUP) إلغاء الإحالة.
10. تعاد كفالة حسن التنفيذ بعد استكمال التوريد وجميع شروط العقد أو أوامر الشراء وبموجب الوثائق الأصولية اللازمة للاستلام.
11. على المشاركين في العطاء ارفاق كتالوجات عن المنتج.
12. يلتزم من يرسو عليه العطاء بدفع غرامة تأخير بواقع (0.1%) عن كل يوم تأخير من قيمة الأعمال المنجزة عن الوقت المحدد في الاتفاقية، ويتم احتساب هذه الغرامات من الدفعات المستحقة له أو من كفالة حسن التنفيذ.
13. يحق لـ (AAUP) إلغاء العطاء دون إبداء الأسباب كما أن (AAUP) غير ملزمة بإحالة العطاء على أقل العروض سعراً دون إبداء الأسباب. ولها أن ترفض كل أو بعض العروض المقدمة لها دون أن يكون لأي من المشاركين الحق في الرجوع إليها بأي خسارة أو ضرر ناجم عن تقديم عرضه ولا يترتب على (AAUP) أي التزامات مادية أو غير مادية مقابل ذلك.



14. يلتزم من يرسو عليه العطاء بتقديم كفالة بنكية (صيانة) بقيمة (5%) من قيمة الأعمال المنجزة صالحة لمدة عام من تاريخ تسليم الأعمال.
15. على المشارك في العطاء تقديم عرضه على أساس المواصفات الفنية المبينة في وثائق العطاء وبموجب الكميات المحددة في جدول الكميات المرفق.
16. لا يجوز للمشارك في العطاء أن يتنازل لأي طرف آخر عن كل أو جزء من أمر الشراء دون الحصول على إذن خطي من (AAUP) مع الاحتفاظ بكامل حقوق (AAUP) وفقاً لشروط أمر الشراء.
17. عند دراسة العروض يؤخذ بعين الاعتبار كفاءة المناقص من الناحيتين المالية والفنية وقدرته على الوفاء بالتزامات العطاء وخبرته في تقديم اللوازم المطلوبة والسمعة التجارية والتسهيلات التي يقدمها ويجوز استبعاد عرضه لنقص كل أو بعض هذه المتطلبات.
18. لا تقبل العروض أو التعديلات التي ترد بعد التاريخ والموعد المحدد كآخر موعد لتقديم العروض.
19. يجب تعبئة جداول المواصفات المرفقة و لن ينظر بأي عرض لا يلتزم بتعبئة الجداول.

- ❖ ويسمح بتقديم عرضان اثنان فقط كحد أقصى لكل بند
- ❖ يجب تقديم عرضي الاسعار الفني والمالي بنسختين: الأولى ورقية، والأخرى الكترونية (محوسية).
- ❖ تقديم العرضين المالي والفني الورقيين بالظرف المختوم، مع ضرورة وضع ختم الشركة والتوقيع على كل الصفحات (للعرض المالي بالذات)



(3)

طريقة الدفع

خلال (90) يوماً من التوريد والقبول والاستلام النهائي، مقابل تقديم الكفالات المطلوبة.



الجزء الثاني

1. جدول الكميات والمواصفات الفنية
عطاء توفير اجهزة مخبرية

| No. | Product | Qty | Unit Price USD | Total Price USD |
|--------------|--|-----|----------------|-----------------|
| 1 | Ultrasonic bath | 1 | | |
| 2 | Mini Centrifuge | 1 | | |
| 3 | Upgrade to inverted microscope Olympus CLX53 | 1 | | |
| 4 | Freeze Dryers (lyophilizes) | 1 | | |
| 5 | FUME HOOD | 2 | | |
| 6 | PCR Cabinet | 1 | | |
| Total | | | | |

في حالة وجود استفسار يرجى تزويدنا بها من خلال البريد الالكتروني للرد عليها pnnp@aaup.edu



1. Ultrasonic bath

- Good Brand name.
- Origin: USA, Europe, japan.
- For cleaning biology lab and research lab
- Microprocessor control .
- Frequency : 37khz .
- With temperature control
- heater up to 85 deg.C
- 3L.
- With cover .
- With basket (immersion basket)
- Anti chemical and corrosive stainless steel .
- With timer

يجب تعبئة الجدول بالأسفل بالإضافة الى تقديم عرض الشركة التقني

| Ultrasonic bath | Compliance Yes / No (if no put your comments) |
|---|--|
| Manufacture | Manufacture |
| Model | Model |
| Origin : USA , Europe | Origin |
| Delivery time | Delivery |
| 3 years warranty | Yes/No |
| For cleaning biology lab and research lab | Yes/No |
| Microprocessor control . | Yes/No |
| Frequency : 37khz . | Yes/No |
| With temperature control | Yes/No |
| 3L. | Yes/No |
| heater up to 85 deg.C | |
| With cover . | Yes/No |
| With basket (immersion basket) | Yes/No |
| Anti chemical and corrosive stainless steel . | Yes/No |
| With timer | Yes/No |
| On/off switch . | Yes/No |
| User's and service manual. | Yes/No |
| Price include installation and team training in site. | Yes/No |
| 3 years warranty, 10 years availability spare parts | Yes/No |
| Date of manufacture not less than 2019 | |



2. Minicentrifuge (minispin)

- Brand name
- Origin : USA , Europe .
- For research lab genetics lab .
- Bench top mini centrifuge with rotor.
- Brushless motor.
- Microprocessor control .
- Max Speed : 2900 RPM .
- Rotor : for microtube and PCR strip with adaptors 0.2 , 0.4 ,0.5 ml
- With Control and display .

❖ Kindly your offer will not be taken into our consideration if you don't fill below table.

| Mini Centrifuge | Yes/No | Comment |
|---|-------------|---------|
| Manufacture | Manufacture | |
| Model | Model | |
| Origin | Origin | |
| Delivery time | weeks | |
| 3 years Full warranty | Yes/No | |
| For research lab genetics lab . | Yes/No | |
| Bench top mini centrifuge with rotor. | Yes/No | |
| Brushless motor. | Yes/No | |
| Microprocessor control . | Yes/No | |
| Max Speed : 2900 RPM . | Yes/No | |
| Rotor : for microtube and PCR strip with adaptors 0.2 , 0.4 ,0.5 ml | Yes/No | |
| With Control and display . | Yes/No | |
| Date of manufacture the machine not less than 2017 | | |
| With quick vibration . | | |
| User's and service manual . | Yes/No | |
| Price include installation and team training in site. | Yes/No | |
| References in Palestine are Attached to your offer | Yes/No | |
| 10 years availability for spare parts | Yes/No | |
| | | |



3. Upgrade to inverted microscope Olympus CLX53

- تطوير ميكروسكوب موجود في الجامعة حسب الموديل المرفق لتحويله الى Fluorescence Microscope
- تلتزم الشركة بتوريد جميع ما يلزم للجهاز .
- تلتزم بتوريد الفلاتر التالية :

- Alexa Fluor® 488 ULTRA Fluorescence Filter Set for LED Light Sources

-Texas Red or Cy3 ULTRA Widefield Fluorescence Filter Set

-DAPI Fluorescence Filter Set for LED Light Sources




4. Freeze Dryers (lyophilizes)

- Good Brand name.
- Origin: USA, Europe, japan.
- Benchtop
- Ice Holding Capacity : 3 Liter .
- Temperature : - 55 C
- Temperature measurement : ice condensation , shelf , product)
- One shelve
- chamber with internal condenser coils, all made from high-grade stainless steel
- Refrigeration system : 0.5kw
- Refrigerant : CFC free
- Noise : 54Db .
- WITH VACUME PUMP 90 LLMIN
- With vacuum control
- With drain Valve

يجب تعبئة الجدول بالأسفل بالإضافة الى تقديم عرض الشركة التقني

| Freeze Dryers (lyophilizes) | Compliance Yes / No (if no put your comments) |
|---|---|
| Manufacture | Manufacture |
| Model | Model |
| Origin : USA , Europe | Origin |
| Delivery time | Delivery |
| 3 years warranty | Yes/No |
| benchtop | Yes/No |
| Ice Holding Capacity : 3 Liter | Yes/No |
| Temperature : - 55 C | Yes/No |
| Temperature measurement : ice condensation , shelf , product) | Yes/No |
| One shelve | Yes/No |
| chamber with internal condenser coils, all made from high-grade stainless steel | Yes/No |
| Refrigeration system : 0.5kw | Yes/No |
| Refrigerant : CFC free | Yes/No |
| Noise : 54Db . | Yes/No |
| WITH VACUME PUMP 90 LLMIN | Yes/No |
| With vacuum control | Yes/No |
| With drain Valve | Yes/No |
| Complete System ready to use with all accessories | Yes/No |
| 220 V | Yes/No |
| User's and service manual. | Yes/No |
| Price include installation and team training in site. | Yes/No |
| 3 years warranty, 10 years availability spare parts | Yes/No |
| Date of manufacture not less than 2019 | Yes/No |



5. FUME HOOD

Brand name

Hood to protects operator from toxic vapor and environment

Application : for general purpose chemistry Lab

Working Dimension : 130 cm w x 70 cm D x 75 cm .

Microprocessor control (clear and simple) .

Construction and Material: Made of good quality melamine board with good acid and alkali resistance function anti chemical, heat, fire resistant and easily cleanable .

Sash : Tempered Glass vertical rising sash.

Sink, Tap Water, Drainage: At least one sink made of porcelain required inside the hood with inlet and outlet arrangement.

With Gas line water line and electric socket (water proof) .

Lighting: fluorescent lamp with vapor proof design .

Blower :

Built-in centrifugal blower for air suction in hood, silent (compatible with above specification) .

Adjustable air speed.

casters and locking for casters .

with stand .

Noise level : 55 db.

Ducting :

- chemical resistant rigid and flexible ductwork from hood to exhaust .
- التركيب يشمل توصيل الجهاز لتمديدات وتركيب في الموقع المخصص

❖ *Kindly your offer will not be taken into our consideration if you don't fill below table.*

| Fume hood | Yes/No | Comment |
|---|---------------|---------|
| Manufacture | Manufacture | |
| Model | Model | |
| Origin | Origin | |
| Delivery time | Delivery time | |
| 3years Full warranty | Yes/No | |
| Hood to protects operator from toxic vapor and environment | Yes/No | |
| Application : for general purpose chemistry Lab | Yes/No | |
| Working Dimension : 130 cm w x 70 cm D x 75 cm . | Yes/No | |
| Microprocessor control (clear and simple) . | Yes/No | |
| Construction and Material: Made of good quality melamine board with good acid and alkali resistance function anti chemical, heat, fire resistant and easily cleanable . | Yes/No | |
| Sash : Tempered Glass vertical rising sash. | Yes/No | |
| Sink, Tap Water, Drainage: At least one sink made of porcelain required inside the hood with inlet and outlet arrangement. | Yes/No | |
| With Gas line water line and electric socket (water proof) . | Yes/No | |
| Lighting: fluorescent lamp with vapor proof design . | Yes/No | |
| Blower : | Yes/No | |



| | | |
|--|--------|--|
| Built-in centrifugal blower for air suction in hood, silent (compatible with above specification) . Adjustable air speed. | | |
| casters and locking for casters . | Yes/No | |
| with stand . | Yes/No | |
| Noise level : 55 db. | | |
| Ducting : | | |
| chemical resistant rigid and flexible ductwork from hood to exhaust . | Yes/No | |
| Should have FDA CE approval , NSF listed | Yes/No | |
| 10 years availability for spare parts | Yes/No | |
| Price include installation and training in site | Yes/No | |
| Complete and Ready to use with all accessories | Yes/No | |
| User's and service manual | Yes/No | |
| Date of manufacture not less than 2019 | Yes/No | |



6. PCR Cabinet

- Good Brand name.
- Origin: USA, Europe, japan.
- interior Dimensions: 60W x 60 D X 60 H cm
- Light : UV , White
- Two power outlets 220 v Type H
- UV Timer .
- Antimicrobial coated metal surfaces .

يجب تعبئة الجدول بالأسفل بالإضافة الى تقديم عرض الشركة التقني

| <u>PCR Cabinet</u> | <u>Compliance Yes / No</u> (if no put your comments) |
|--|---|
| <u>Manufacture</u> | Manufacture |
| <u>Model</u> | Model |
| <u>Origin : USA , Europe</u> | Origin |
| <u>Delivery time</u> | Delivery |
| <u>3 years warranty</u> | Yes/No |
| <u>interior Dimensions: 60W x 60 D X 60 H cm</u> | Yes/No |
| <u>Light : UV , White</u> | |
| <u>Two power outlets 220 v Type H</u> | |
| <u>UV Timer .</u> | |
| <u>Antimicrobial coated metal surfaces .</u> | |
| <u>User's and service manual.</u> | Yes/No |
| <u>Price include installation and team training in site.</u> | Yes/No |
| <u>3 years warranty, 10 years availability spare parts</u> | Yes/No |
| <u>Date of manufacture not less than 2019</u> | |

