

### الجامعة العربية الإمريكية دائرة اللوازم والمشتريات

		13/8/2024	لتاربخ:
R-Q107-2023.24	رقم الاستدراج:		
			لسادة: .
			لفون:
			يميل:

R-BR2324-000

استدراج عروض أسعار: الرجاء تزويدنا بأسعار اللوازم التالية شامل لضريبة القيمة المضافة:

اجمالي تكلفة البند	السعر	الكمية	الوحدة	اسم الصنف	
		1.00	37E	Defibrillator Machine جهاز صدمات القلب مرفق المواصفات الفنية	1
	شيقل	کلي (ش.ض)	المجموع الدُ		

ختم وتوقيع المورد:	تعليمات الاستدراج		
	الأربعاء	آخر موعد لتسليم العروض هو يوم: **	
	21/8/2024	** الموافق:	
	جهاز صدمات القلب	** طبيعة البنود المطلوبة:	
	شيقل	** بعملة:	
	التوصيل الى حرم رام الله	** الأسعار تشمل التوصيل الى	
	بالظرف المغلق	** يرجى إرسال العرض ب:	

### شروط الاستدراج:

- \*\* ضرورة تحديد مدة الكفالة إن وجدت.
- \*\* ضرورة تحديد مدة التوريد من تاريخ استلام امر الشراء، و في حال عدم الالتزام يتم فرض غرامة تأخير .
- \*\* ختم وتوقيع المورد المعتمد على هذا النموذج، وكتابة رقم الاستدراج واسم الشركة كما هو مسجل لدى الضريبة على الظرف.
  - \*\* للمراجعة والاستفسار بخصوص المواصفات يرجى الاتصال على هاتف 022941999 رقم داخلي1910
    - \*\* للمراجعة والاستفسار بخصوص المواصفات عن طريق بريد الكتروني Rasha.kmail@aaup.edu
    - \*\* لمشاهدة العطاءات والاستدراجات، يرجى الدخول الى الرابط http://www.aaup.edu/tenders





# الجامعة العربية الأمريكية

### **Defibrillator Machine**

#### Brand name.

3 years warranty, 10 years availability of spare parts.

Origin: USA, EURO, Japan.

### **General Specifications:**

- Designed for use in medical college teaching nursing.
- Incorporates defibrillation, non-invasive pacing, and monitoring capabilities.
- Equipped with CO2 mainstream capnograph measurement for comprehensive patient assessment.

#### **Defibrillation:**

- Utilizes biphasic waveform for effective defibrillation.
- Impedance range: 25  $\Omega$  to 200  $\Omega$  for external defibrillation.
- Energy selections range from 1 J to 360 J, adjustable in increments.

### **Charge Time:**

- Reaches 200 J in less than 5 seconds with a new, fully charged battery.
- Reaches 360 J in less than 8 seconds with a new, fully charged battery.

#### Modes:

- Manual mode: supports asynchronous and synchronous defibrillation.
- AED mode: enables automated external defibrillation for rapid response.

### **Non-Invasive Pacing:**

- Offers fixed and on-demand pacing modes.
- Delivers monophasic square wave pulse for effective pacing.

#### Monitoring:

- Provides three curves, including 5-lead ECG, SpO2, non-invasive blood pressure, respiration, EtCO2 (CO2 mainstream capnograph measurement), temperature (T1 & T2), and events monitoring.
- Enables comprehensive patient assessment and continuous monitoring during treatment.

### Standards:

- Complies with relevant medical device standards for safety and performance.
- Ensures reliable operation and accurate monitoring in clinical settings.

### Additional Features:

tiser triendly interface with intuitive controls for easy operation.

Lightweight and portable design for flexible use in various healthcare settings.

ائرة اللهازم والمشتريبات - رام الله PROCUREMENT AND PURCHASIN DEPARTMENT - RAMALLAH

SAS AMERICAN UNIVERSITY ARAB AMERICAN UNIVERSITAND



# الجامعة العربية الأمريكية

- Long battery life and quick recharge capability for continuous operation during emergencies. **Support:**
- Offers comprehensive technical support and training for users.
- Availability of spare parts and service technicians for maintenance and repair needs. Please price separately all accessories and consumables

السعر يشمل التوريد والتركيب ويجب أن يتم توضيح جميع متطلبات التركيب في عرض السعر من ملحقات ومتطلبات. للسعر يشمل التوريد والتركيب ويجب أن يتم توضيح جميع متطلبات التركيب في عرض السعر من ملحقات ومتطلبات.

<u>Defibrillator Machine</u>	Yes/No	Comment
Manufacture	Manufacture	
Model	Model	
Origin: USA, EUPRE, japan.	Origin	
Delivery time	Weeks	
3 years Full warranty 10 years availability	Yes/No	
for spare parts		
General Specifications:		
Designed for use in medical college teaching nursing.	Yes/No	
Incorporates defibrillation, non-invasive	Yes/No	
pacing, and monitoring capabilities.		
Equipped with CO2 mainstream capnograph	Yes/No	
measurement for comprehensive patient		
assessment.		
Defibrillation:		
Utilizes biphasic waveform for effective defibrillation.	Yes/No	
Impedance range: 25 $\Omega$ to 200 $\Omega$ for external defibrillation.	Yes/No	
Energy selections range from 1 J to 360 J, adjustable in increments.	Yes/No	
Charge Time:		
Reaches 200 J in less than 5 seconds with a new, fully charged battery.	Yes/No	
Reaches 360 J in less than 8 seconds with a new, fully charged batter.	Yes/No	
Modes:		

PROCUREMENT AND PURCHASIN
DEPARTMENT - RAMALLAH

OFFICIAL PROCURE OF THE PROCURE

pg. 2



# الجامعة العربية الأمريكية

Manual modes summerts asymphesis and	Yes/No	
Manual mode: supports asynchronous and	N CN/IND	
synchronous defibrillation.  AED mode: enables automated external	3773.7	
	Yes/No	
defibrillation for rapid response.		
Non-Invasive Pacing:	W. T. 1917	
Offers fixed and on-demand pacing modes.	Yes/No	
Delivers monophasic square wave pulse for	Yes/No	
effective pacing.		
Monitoring:		
Provides three curves, including 5-lead ECG,	Yes/No	
SpO2, non-invasive blood pressure,		
respiration, EtCO2 (CO2 mainstream		
capnograph measurement), temperature (T1		
& T2), and events monitoring.		
Enables comprehensive patient assessment	Yes/No	
and continuous monitoring during treatment.		
Standards:		<u> </u>
Complies with relevant medical device	Yes/No	
standards for safety and performance.		
Ensures reliable operation and accurate	Yes/No	
monitoring in clinical settings.		
Additional Features:		
User-friendly interface with intuitive controls	Yes/No	
for easy operation.		
Lightweight and portable design for flexible		
use in various healthcare settings.		
Long battery life and quick recharge	Yes/No	
capability for continuous operation during		
emergencies.		
Support:		
Offers comprehensive technical support and	Yes/No	
training for users.		
Availability of spare parts and service	Yes/No	
technicians for maintenance and repair needs.	The same and the same	
, , , , , , , , , , , , , , , , , , ,		
Please price separately all accessories and	Yes/No	
consumables.		
The price includes installation and team	Yes/No	
training on-site.		
Date of manufacture of the machine not	Yes/No	
less than 2024		





# الجامعة العربية الأمريكية

Your warranty must include all spare parts and labor.	Yes/No	
Your warranty includes at least two visits at least (preventive maintenance) onsite for each year's warranty from the installation date.	Yes/No	
10 years Warranty for spare parts Availability from date of instillation	Yes/No	
Service & and operation manuals and spare parts list are included	Yes/No	

