

التاريخ: 04/01/2021

رقم الاستدراج: Q136-2022

السادة:

تلفون:

إيميل:

BR2122-000448

استدراج عروض أسعار: الرجاء تزويدنا بأسعار اللوازم التالية شامل لضريبة القيمة المضافة:

اسم الصنف	الوحدة	الكمية	السعر	اجمالي تكلفة البند
Medical O2 generator المواصفات صفحة 2 و3	عدد	1		
Vein Illumination system المواصفات صفحة 4 و5	عدد	1		
المجموع الكلي (ش.ض)			دولار	

تعليمات الاستدراج	ختم وتوقيع المورد:
** آخر موعد لتسليم العروض هو يوم:	الثلاثاء
** الموافق:	11/01/2022
** طبيعة البنود المطلوبة:	أجهزة عيادة طبية
** بعملة:	دولار
** الأسعار تشمل التوصيل ل:	الحرم الرئيسي - جنين
** نوع التوريد:	توريد كامل
** يرجى إرسال العرض ب:	بالإيميل

شروط الاستدراج:
** ضرورة تحديد مدة الكفالة إن وجدت.
** ضرورة تحديد مدة التوريد من تاريخ استلام امر الشراء، وفي حال عدم الالتزام يتم فرض غرامة تأخير .
** ختم وتوقيع المورد المعتمد على هذا النموذج، وكتابة رقم الاستدراج واسم الشركة كما هو مسجل لدى الضريبة على الظرف.
** للمراجعة والاستفسار بخصوص المواصفات يرجى الاتصال على هاتف 042418888 رقم داخلي 1481
** للمراجعة والاستفسار بخصوص المواصفات عن طريق فاكس 042510972 أو بريد الكتروني manal.daraghmah@aaup
** لمشاهدة العطاءات والاستدراجات، يرجى الدخول الى الرابط http://www.aaup.edu/tenders

مدير دائرة اللوازم والمشتريات



5/1

Item 1

Medical O2 generator

Medical o2 generator
easy adjustable
with flow meter
24hours / 7 days continuous suppling o2 .
Fault diagnostic and safety
with alarm system at low o2 concentration .
Flow rate 10L/min
o2 purity : $\geq 93 \pm 3\%$
Noise : ≤ 50 dB
Nasal Cannula, adult mask, humidifier bottle, Oxygen tube, Nebulizer cup
Portable with casters .

يجب تعبئة الجدول بالأسفل بالإضافة الى تقديم عرض الشركة التقني

<u>Medical O2 generator</u>	Compliance Yes / No (if no put your comments)
Manufacture	Manufacture
Model	Model
Origin	Origin
Delivery time	Delivery time
3 years warranty, 10 years availability spare parts.	Yes / No
Medical o2 generator	Yes / No
easy adjustable	Yes / No
with flow meter	Yes / No
24hours / 7 days continuous suppling o2 .	Yes / No
Fault diagnostic and safety	Yes / No
with alarm system at low o2 concentration .	Yes / No
Flow rate 10L/min	Yes / No
o2 purity : $\geq 93 \pm 3\%$	Yes / No
Noise : ≤ 50 dB	Yes / No
Nasal Cannula, adult mask, humidifier bottle, Oxygen tube, Nebulizer cup	Yes / No
Portable with casters .	Yes / No
24hours / 7 days continuous suppling o2 .	Yes / No



Fault diagnostic and safety	Yes / No
with alarm system at low o2 concentration .	Yes / No
Your warranty must be include all spare parts and labor	Yes / No
Price should be include installation and team training in site .	Yes / No
Service & operation manuals are included & Date of manufacturing of system not less than 2022	Yes / No

5/3



Item 2

Vein Illumination system

3years warranty
To identifying Vein
Vein illumination device
Without calibration and ready to use
Different zoom and color
Adult pediatric and dark skin .
Handheld
Non contact
Lightweight
Easy to use
With rechargeable battery and charger working up to 3 hours
With mobile trolley

يجب تعبئة الجدول بالأسفل بالإضافة الى تقديم عرض الشركة التقني

<u>Vein Illumination system</u>	Compliance Yes / No (if no put your comments)
Manufacture	Manufacture
Model	Model
Origin	Origin
Delivery time	Delivery time
2 years warranty, 10 years availability spare parts.	Yes / No
To identifying Vein	Yes / No
Vein illumination device	Yes / No
Without calibration and ready to use	Yes / No
Adult pediatric and dark skin . Different zoom and color	Yes / No
Handheld	Yes / No
Non contact	Yes / No
Lightweight	Yes / No
Easy to use	Yes / No

514



With rechargeable battery and charger working up to 3 hours	Yes / No
Price should be include installation and team training in site .	Yes / No
Service & operation manuals are included & Date of manufacturing of system not less than 2022 .	Yes / No
With mobile trolley	Yes / No

5/5

