

التاريخ: 14/3/2018

السادة المحترمين

تلفون:

فاكس:

(Q196-2017.18)

رقم : الاستدراج
/الشراء المباشر/الشراء بالتزيم

DEFIBRILLATOR with pacing

الموضوع : عروض اسعار

إشارة إلى الموضوع اعلاه الرجاء تزويدنا بأسعار اللوازم التالية (بالدولار) شامل لضريبة القيمة المضافة والتوصيل للجامعة العربية الأمريكية:

| الرقم | اسم الصنف | الوحدة | الكمية | السعر | اجمالي تكلفة البند |
|-------|---------------------------|--------|--------|---------------|--------------------|
| 1 | DEFIBRILLATOR with pacing | Each | 1 | | |
| | | | | المجموع الكلي | |

شروط الاستدراج

- 1- يرجى العلم ان اخر موعد لتسليم العروض هو يوم (الأربعاء) الموافق 20/3/2018
- 2- الاسعار شاملة لضريبة القيمة المضافة.
- 3- ضرورة تحديد مدة التوريد من تاريخ استلام امر الشراء، و في حال عدم الالتزام يتم فرض غرامة تأخير .
- 4- ختم وتوقيع المورد المعتمد على هذا النموذج، وكتابة رقم الاستدراج واسم الشركة كما هو مسجل لدى الضريبة على الطرف.
- 5- لا تقبل العروض الا بالطرف المختوم.
- 6- للمراجعة والاستفسار بخصوص المواصفات يرجى الاتصال على هاتف 042418888 رقم داخلي 1481 او فاكس 042510972 أو بريد الكتروني manal.daraghmah@auj.edu
- 7- لمشاهدة العطاءات والاستدراجات، يرجى الدخول الى الرابط <http://www.auj.edu/Tenders> مع الاحترام،

مدير دائرة اللوازم والمشتريات

أ. حسن ربيعة

14/3/2018

1/1

DEFIBRILLATOR with pacing

- Origin: USA , japan , Europe .
- Biphasic defibrillator with pacing .
- Clear display.
- Automatic lead switching to see patient ECG through paddles or leads .
- Built in 50mm strip printer/ thermal recorder.
- Charging time of less than 3 seconds for maximum energy. Charging indicator should be there.
- Bright electroluminescent display for viewing messages and ECG waveform of 4 seconds.
- External & internal paddles with paddles contact indicator (with adult pediatric paddles)
- event summary facility for recording and printing at least 250 events and 50 waveforms.
- battery capable of usage for at least 90minutes or 30 discharges.
- Should be capable of printing Reports on Event summary, configuration, self-test, battery capacity etc.
- facility for self-test/check before usage and set up function.
- Should be capable of delivering energy in increments of 1-2

Optional price

- Paddles adult.
- Paddles pediatric
- ECG Leads.

يجب تعبئة الجدول بالأسفل بالإضافة الى تقديم عرض الشركة التقنى

| DEFIBRILLATOR | Compliance Yes / No (if no put your comments) |
|---|--|
| Manufacture | Manufacture |
| Model | Model |
| Origin | Origin |
| Delivery time | Delivery |
| 3 years warranty | |
| Biphasic defibrillator, with pacing . | Yes/No |
| Clear display . | Yes/No |
| Automatic Lead switching to see patient ECG through paddles or leads. | Yes/No |
| built in 50mm strip printer/ thermal recorder. | Yes/No |
| charging less than 3s for maximum energy. Charging indicator should be there. | Yes/No |
| bright electroluminescent display for viewing messages and ECG waveform of 4 seconds. | Yes/No |
| external & internal paddles with paddles contact indicator | Yes/No |
| event summary facility for recording and printing at least 250 events and 50 waveforms. | Yes/No |
| battery capable of usage for at least 90minutes or 30 discharges. | Yes/No |
| Should be capable of printing Reports on Event summary, configuration, self-test, battery capacity etc. | Yes/No |
| facility for self-test/check before usage and set up function. | Yes/No |
| capable of delivering energy in increments of 1-2 | Yes/No |
| User's and service manual . | Yes/No |
| Price include installation and team training in site. | Yes/No |
| 10 years availability for spare parts | Yes/No |
| Date of manufacture of system not less than 2018 | |

2/2



 14/13/2018