

التاريخ: 10/4/2018

السادة المحترمين

تلفون:

فاكس:

(Q226 -2017.18)

رقم : الاستدراج /
الشراء المباشر/الشراء بالتزيم

الموضوع : عروض اسعار Nano photometer

إشارة إلى الموضوع اعلاه الرجاء تزويدنا بأسعار اللوازم التالية (بالدولار) شامل لضريبة القيمة المضافة والتوصيل للجامعة العربية الأمريكية:

الرقم	اسم الصنف	الوحدة	الكمية	السعر	اجمالي تكلفة البند	ملاحظات
1	جهاز Nano photometer المواصفات الفنية مرفقة ورقة 3.2	EACH	1			
				المجموع الكلي		

شروط الاستدراج

- 1- يرجى العلم ان اخر موعد لتسليم العروض هو يوم (الاحد) الموافق 22/4/2018
- 2- الاسعار شاملة لضريبة القيمة المضافة.
- 3- ضرورة تحديد مدة التوريد من تاريخ استلام امر الشراء، و في حال عدم الالتزام يتم فرض غرامة تأخير .
- 4- ختم وتوقيع المورد المعتمد على هذا النموذج، وكتابة رقم الاستدراج واسم الشركة كما هو مسجل لدى الضريبة على الظرف.
- 5- لا تقبل العروض الا بالظرف المختوم.
- 6- للمراجعة والاستفسار بخصوص المواصفات يرجى الاتصال على هاتف 042418888 رقم داخلي 1491 او فاكس 042510972 أو بريد الكتروني manal.daraghmah@aaui.edu@
- 7- لمشاهدة العطاءات والاستدراجات، يرجى الدخول الى الرابط <http://www.aaui.edu/Tenders>

مع الاحترام،

مدير دائرة اللوازم والمشتريات
أ. حسن ربيعة




3/1

Nano photometer

- Application : life science lab .
- UV Vis spectrophotometer.
- Wavelength range: 190 - 800 nm .
- Wavelength accuracy: 1 nm .
- Measurement Time : ≤ 5 s .
- Min. Sample volume: ≤ 0.5 ul .
- Lamp : Xenon flash lamp .
- Sampling mode : dual with microliter volume .
- Lowest Detection range: 2 ng/ μ l for dsDNA , 0.01 mg/ml BSA
- Highest range : not less than 15000 ng/ μ l for dsDNA , 400 mg/ml BSA .
- Dust cover .
- Main price should be Completed and ready to use with all stander accessories without any addition cost .

3/2



❖ Kindly your offer will not be taken into our consideration if you don't fill below table.

<u>Nano photometer</u>	Yes/No	Comment
Manufacture	Manufacture	
Model	Model	
Origin : Europe , USA .	Origin	
Delivery time	Weeks	
<u>3 years Full warranty</u>	Yes/No	
Brand name	Yes/No	
Application : life science lab .	Yes/No	
UV Vis spectrophotometer.	Yes/No	
Wavelength range: 190 - 800 nm .	Yes/No	
Wavelength accuracy: 1 nm .	Yes/No	
Measurement Time : ≤ 5 s .	Yes/No	
Min. Sample volume: ≤ 0.5 ul .	Yes/No	
Lamp : Xenon flash lamp .	Yes/No	
Sampling mode : dual with microliter volume .	Yes/No	
Lowest Detection range: 2 ng/ μ l for dsDNA , 0.01 mg/ml BSA	Yes/No	
Highest range : not less than 15000 ng/ μ l for dsDNA , 400 mg/ml BSA	Yes/No	
Dust cover .	Yes/No	
Main price should be Completed and ready to use with all stander accessories without any addition cost	Yes/No	
Authorization letter from the manufacture to represent them	Yes/No	
User's and service manual .	Yes/No	
Price include installation and team training in site.	Yes/No	
10 years availability for spare parts	Yes/No	
Date of manufacture the machine not less than 2018	Yes/No	

3/3

10.11.2018

