

التاريخ: 03/02/2020

السادة: المحترمين
تلفون:
فاكس:

رقم / الاستدراج (Q149/2019.20)

جهاز طبي Vein Illumination system

الموضوع: عروض اسعار

إشارة إلى الموضوع اعلاه الرجاء تزويدنا بأسعار اللوازم التالية (بالدولار) شامل لضريبة القيمة المضافة والتوصيل للجامعة العربية الأمريكية:

#	اسم الصنف	الوحدة	الكمية	السعر	اجمالي تكلفة البند
2	Vein Illumination system مرفق المواصفات على الصفحة رقم 2	عدد	1		
				المجموع الكلي	

شروط الاستدراج

- 1- يرجى العلم ان اخر موعد لتسليم العروض هو يوم الإثنين الموافق: 10/02/2020
- 2- الاسعار شاملة لضريبة القيمة المضافة.
- 3- ضرورة تحديد مدة التوريد من تاريخ استلام امر الشراء، و في حال عدم الالتزام يتم فرض غرامة تأخير .
- 4- ختم وتوقيع المورد المعتمد على هذا النموذج، وكتابة رقم الاستدراج واسم الشركة كما هو مسجل لدى الضريبة على الظرف.
- 5- لا تقبل العروض الا بالظرف المختوم.
- 6- للمراجعة والاستفسار بخصوص المواصفات يرجى الاتصال على هاتف 042418888 رقم داخلي 1481 او فاكس 042510972 أو بريد الكتروني manal.daraghmah@aup.edu
- 7- لمشاهدة العطاءات والاستدراجات، يرجى الدخول الى الرابط <http://www.aup.edu/Tenders>



2/1

Vein Illumination system

2 years warranty
To identifying Vein
Vein illumination device
Without calibration and ready to use
Adult pediatric and dark skin .
Handheld
Non contact
Lightweight
Easy to use
With rechargeable battery working up to 3 hours

Optional :
With mobile

يجب تعبئة الجدول بالأسفل بالإضافة الى تقديم عرض الشركة التقني

<u>Vein Illumination system</u>	Compliance Yes / No (if no put your comments)
Manufacture	Manufacture
Model	Model
Origin	Origin
Delivery time	Delivery time
2 years warranty, 10 years availability spare parts.	Yes / No
To identifying Vein	Yes / No
Vein illumination device	Yes / No
Without calibration and ready to use	Yes / No
Adult pediatric and dark skin .	Yes / No
Handheld	Yes / No
Non contact	Yes / No
Lightweight	Yes / No
Easy to use	Yes / No
With rechargeable battery working up to 3 hours	Yes / No
<u>Optional : with mobile</u>	
Price should be include installation and team training in site .	Yes / No
Service & operation manuals are included & Date of manufacturing of system not less than 2020 .	Yes / No
	Yes / No

2/2

